



Angiogenesis Co-Culture Assay Kit Protocol (GFP-HUVEC/HDF)

Product Code: ZHA-4100-96

The cells in this kit require immediate attention

For research use only – not for diagnostic or therapeutic use

**Please read the entire protocol prior to use
and take care to follow the instructions carefully**

Kit Contents

Room temperature box contents:

- Angiogenesis Co-Culture Assay Kit Protocol (GFP-HUVEC/HDF)
- Certificate of Conformity
- KC1000 96 well tissue culture plate
- KC1003 Angiogenesis basal medium (125ml)

Dry ice box contents:

- KC1032 Angiogenesis co-culture cells (GFP-HUVEC/HDF)
- KC1036 Growth supplement (2.5ml)
- KC1016 Antibiotic supplement (125µl)
- KC1006 VEGF 2ng/ml (20µl)
- KC1007 Suramin 1mM (220µl)



A Caltag Medsystems Company

Contents

1. Introduction
2. Handling and Storage of Kit Components
3. Additional Materials
4. Guidelines for Test Conditions
5. Protocol Summary
6. Angiogenesis Co-Culture Assay Protocol
7. Data Analysis
8. Troubleshooting
9. Related Cellworks Products
10. References
11. Technical Support

1. Introduction

Angiogenesis is the multistep process whereby new blood vessels develop from pre-existing vasculature. Angiogenesis plays a key role in numerous physiological and pathological processes including wound healing and the development of collateral circulation following an ischaemic episode¹, reproduction-associated neovascularisation², growth of solid tumours³ and diabetic retinopathy⁴. Understanding the mechanism of angiogenesis will therefore provide new approaches to the treatment of a wide range of pathologies.

Angiogenesis is a complex process in which the following events are believed to play a critical role:

- Proteolytic degradation of the extracellular matrix⁵
- Directed migration of endothelial cells^{6,7}
- Proliferation of endothelial cells⁸
- Deposition of new extracellular matrix⁵
- Formation of tubules and anastomosis of the newly formed vessels^{5,7}

Experimental approaches to the study of these events have been limited by the lack of suitable models of angiogenesis. Several *in vivo* systems have been developed including the chick chorioallantoic membrane (CAM) assay⁹ and the rabbit cornea model¹⁰ but these systems are impractical for the study or screening of large numbers of samples and are far removed from angiogenesis in a human system. The *in vitro* methods currently in use have generally isolated the different component parts of the angiogenic process and have studied endothelial cell proliferation¹¹, endothelial cell migration¹² or the ability of endothelial cells to associate into tubules when in contact with various matrix proteins¹³. None of these assay systems accurately reflect the angiogenic process in its entirety.

In the Cellworks Angiogenesis Co-Culture Assay Kit, TagGFP2-expressing human endothelial cells are co-cultured with human fibroblasts in a specially formulated medium. The endothelial cells initially form small islands within the culture matrix. They subsequently begin to proliferate and then enter a migratory phase during which they move through the matrix to form threadlike tubule structures. These gradually join up to form a network of anastomosing tubules which closely resembles the capillary bed found in the CAM assay¹⁴.

The angiogenesis co-culture is responsive to known micro- and macro-molecular inhibitors and stimulators of angiogenesis and so, unlike some other models, can be used to measure both positive and negative effects on angiogenesis. It yields reproducible dose response curves permitting comparison of different treatment regimens and reagent concentrations.

In addition to cryogenically-preserved co-culture cells, a 96 well tissue culture plate, growth medium and associated supplements, the kit also includes:

- Validated positive control (VEGF)
- Validated negative control (Suramin)

The kit is designed so that test compounds or conditioned media can be added to the culture at any time during the angiogenic process. HUVEC TagGFP2 stable expression enables detection of fluorescence signal to measure tubule formation over time and/or at assay end point in each condition. The excitation and emission peaks for TagGFP2 are 483nm and 506nm, respectively. Parameters of tubule networks can be measured from images of the plate using Cellworks Image Analysis Software, AngioSys 2.0. A demonstration version of this software is available free of charge (see section 'Analysing Results' for details).

2. Handling and Storage of Kit Components

Code	Component	Storage Temperature
KC1000	96 well tissue culture plate	Ambient
KC1003	Angiogenesis basal medium	2-8°C
KC1032	Angiogenesis co-culture cells (GFP-HUVEC/HDF)	Between -135 and -196°C (vapour-/liquid-phase nitrogen)
KC1036	Growth supplement	-20°C
KC1016	Antibiotic supplement	-20°C
KC1006	VEGF control	-20°C
KC1007	Suramin control	-20°C

Ensure kit components are stored at the indicated temperatures immediately and are used prior to expiry dates elapsing

Contact techsupport@caltagmedsystems.co.uk for kit component MSDS

BIOHAZARD NOTE

This kit contains cells of human origin that test negative for HIV-1, hepatitis B, hepatitis C, mycoplasma, bacteria and fungi. No test procedure can guarantee the absence of known and unknown pathogenic agents. Human cells should be considered potentially biohazardous and appropriate precautions taken for their handling and disposal.

Implement good laboratory practice and use aseptic technique at all times.

3. Additional Materials

The following equipment and reagents are required in addition to the kit contents

- Class 2 laminar flow hood
- Incubator at 37°C with 5% CO₂ humidified atmosphere
- Aspirator and sterile aspirator pipettes
- Micropipettes and sterile pipette tips
- Electronic power pipette pump and sterile serological pipettes
- Sterile tubes
- Test compounds etc. for test conditions
- Microscope for fluorescence detection (ex/em 483/506)

4. Guidelines for Test Conditions

Test Compounds

Test compounds should be added directly to angiogenesis growth medium whenever possible. If necessary, compounds may be dissolved or reconstituted in a solvent such as DMSO or ethanol. A concentrated stock solution can then be diluted in angiogenesis growth medium to the required working concentration. Final solvent concentrations should be kept constant and the experimental design should include control wells treated with solvent alone. Final solvent concentrations should not exceed 0.1% (v/v).

Conditioned Media

Conditioned cell culture media can be diluted in angiogenesis growth medium and added directly to the plate. Optimal dilution ratios should be determined by the end-user but as a starting point it is recommended that conditioned media are diluted 1:1 in angiogenesis growth medium.

5. Protocol Summary

The following provides an outline of the angiogenesis co-culture assay kit protocol

We recommend at least six replicate wells per control and test condition

Day 1

- Angiogenesis growth medium preparation
- Angiogenesis co-culture cell seeding

Day 2

- Control and test compound dilutions (as required)
- Medium change

Days 3 to 10

- Medium change every 2 to 3 days

Day 11

- Assay endpoint

Tubules can be imaged throughout the assay and/or at the recommended assay endpoint of day 11

6. Angiogenesis Co-Culture Assay Protocol

The following provides details of the angiogenesis co-culture assay kit protocol including use of control and test compounds. This protocol can be adapted for an end-user's specific requirements, for example testing conditioned media.

Use good laboratory practice and aseptic technique at all times

Day 1 – assay set-up

- Thaw growth supplement and antibiotic supplement at 2-8°C (overnight) or at room temperature
- Equilibrate angiogenesis basal medium to room temperature
- Add growth supplement and antibiotic supplement to angiogenesis basal medium and mix well to formulate angiogenesis growth medium (AGM)
Omission of antibiotic supplement may lead to excessive tubule formation
- Aliquot 24ml AGM into a sterile tube
- Rapidly thaw the co-culture cells by agitating the vial in a 37°C water bath
- Immediately add 650µl cells to the AGM aliquot
Avoid adding the entire 1ml vial contents to the AGM aliquot
- Mix gently using a serological pipette or by inverting the tube
- Add 150µl cell suspension to each well of the 96 well plate, ensuring the cell suspension remains evenly mixed during this process
- Ensure even dispersion of the cells within each well by holding the plate horizontally and rotating clockwise, anticlockwise and in a figure-of-eight several times
This is important to reduce cell clumping and uneven distribution of tubules
- Place the plate on the bench for 20-30 min to allow the cells to adhere
Avoid leaving the cells at room temperature for more than 30 min
- Place the plate in an incubator at 37°C/5% CO₂ humidified atmosphere

Day 2 – addition of compounds

- Equilibrate AGM to room temperature
- Thaw control compounds at room temperature
- Dilute control and test compounds as follows:
 - VEGF 1:1000 (e.g. add 11µl VEGF solution to 10.989ml AGM)
 - Suramin 1:50 (e.g. add 220µl suramin solution to 10.78ml AGM)
 - Aliquot 11ml AGM (or equivalent volume to control solutions) for untreated control conditions
 - Dilute any test compounds as required*Prepare adequate volumes for all medium changes for the required number of wells for the duration of the assay (11ml is ample for four medium changes for up to 16 wells per condition)*
- Change the medium for up to eight wells at a time:
 - Aspirate the medium from each well, taking care to avoid contact with the cells
 - Gently add 150µl of the required equilibrated medium down the side of each well
Avoid possible cell desiccation by changing the medium for less than eight wells at a time
Do not mix test compounds directly in the wells as this will likely damage the cell culture
- Return the plate to the incubator at 37°C/5% CO₂ humidified atmosphere
- Store medium aliquots at 2-8°C when not in use

Days 3 to 10 – medium changes every 2 to 3 days

- Equilibrate medium aliquots to room temperature
- Change the medium for up to eight wells at a time as Day 2
- Return the plate to the incubator at 37°C/5% CO₂ humidified atmosphere

7. Data Analysis

Appropriate methods of imaging and analysis should be determined by the end-user. Below is a brief description of the Cellworks software that can be used for tubule formation quantification with ImageJ-processed image files.

AngioSys 2.0

Cellworks Image Analysis Software, AngioSys 2.0 (Product Code ZHA-5000), can be used for semi-automated analysis of tubule networks. Measurable parameters include tubule number, tubule branching (number of junctions) and tubule length (mean and total tubule length per image).

Image files can be conveniently grouped and processed to provide quantitative and repeatable measurements. Repetitive image processing sequences can be specified, saved and applied to groups of images to enable rapid analysis. The resulting data is saved in a text format that can easily be read by third party software such as Microsoft Excel.

Further information including details of a free demonstration version can be viewed at:

https://www.cellworks.co.uk/angiogenesis_image_analysis_software.php

8. Troubleshooting

Issues	Causes/Solutions
<p>Poor cell growth and/or little to no tubule formation</p>	<p>Inconsistent cell culture conditions – ensure the incubator is maintained at 37°C/5% CO₂ humidified atmosphere throughout the assay</p> <p>Incorrect storage or formulation of medium (AGM) – ensure medium components are stored at the appropriate temperatures, added as described and that AGM is stored at 2-8°C (do not freeze)</p> <p>Cell desiccation upon change of medium – replace medium in fewer wells at a time</p> <p>Toxic test compound – prepare stock solution of test compound at a lower concentration</p> <p>Toxic concentration of solvent in test compound conditions – prepare stock solutions of test compounds at a higher concentration and include solvent only control wells</p>
<p>Floating cells and/or cell debris</p>	<p>A small number of floating cells and/or a small amount of debris is typical of cell culture</p> <p>Failure of cells to adhere – inconsistent cell culture conditions or incorrect storage or formulation of medium (as above)</p> <p>Accumulation of floating cells and/or debris during the assay – inconsistent cell culture conditions, cell desiccation or toxic test conditions (as above)</p>
<p>Cell sheet detachment</p>	<p>Inadequate care taken when aspirating and/or dispensing – avoid contacting the well base when aspirating and be sure to gently dispense medium or reagent down the side of each well (cell sheet disruption is more likely with use of a single or multi-channel pipette without due care.</p> <p>Uneven dispersion of cells upon seeding, leading to uneven cell sheet thickness – after adding cells to the wells, rotate the plate as described and incubate on the bench for 20-30 min before placing in the incubator</p>
<p>Contamination of one or more wells</p>	<p>Contamination arises from inadequate sterile technique and is more likely to occur if the antibiotic/antimycotic supplement is omitted from the medium</p> <p>Contain contamination by treating affected wells with 1M NaOH for 2-3 hours then aspirating and leaving empty for the remainder of the assay</p>
<p>Appearance of HUVEC 'islands'</p>	<p>Uneven dispersion of cells upon seeding, leading to HUVEC clumps – after adding cells, rotate the plate as described and incubate on the bench for 20-30 min before placing in the incubator</p>

9. Related Cellworks Products

Product type	Product code	Name	Description
Cellworks Cells	ZHC-2301	Early Passage Human Umbilical Vein Endothelial Cells (HUVEC)	Passage 2, pooled donor, cryopreserved, 0.5×10^6 cells/ml/vial
	ZHC-2102	HUVEC (angiogenesis co-culture validated)	Passage 2, pooled donor, cryopreserved, 0.5×10^6 cells/ml/vial
	ZHC-2402	GFP-HUVEC (angiogenesis co-culture validated)	Stably expressing TagGFP2, passage 2, pooled donor, cryopreserved, 0.5×10^6 cells/ml/vial
	ZHC-5102	Human Dermal Fibroblasts (HDF) (angiogenesis co-culture validated)	Passage 8, single adult donor, cryopreserved, 1×10^6 cells/ml/vial
	ZHC-3311	Early Passage Human Coronary Artery Smooth Muscle Cells (HCASMC)	Passage 4, single adult donor, cryopreserved, 0.5×10^6 cells/ml/vial
AngioCo Angiogenesis Co-Culture Assay Kits	ZHA-4000-24	Angiogenesis Co-Culture Assay (HUVEC/HDF) 24wp	Angiogenesis assay kit for immunohistochemical staining of tubules (24 well plate format)
	ZHA-4100-24	Angiogenesis Co-Culture Assay (GFP-HUVEC/HDF) 24wp	Angiogenesis assay kit for green fluorescent detection of tubules (TagGFP2-expressing HUVEC) (24 well plate format)
	ZHA-4100-96	Angiogenesis Co-Culture Assay (GFP-HUVEC/HDF) 96wp	Angiogenesis assay kit for green fluorescent detection of tubules (TagGFP2-expressing HUVEC) (96 well plate format)
AngioCo Angiogenesis Co-Culture Assay Reagents	ZHA-1300	Angiogenesis Control Reagent Kit	VEGF positive control and Suramin negative control
	ZHA-1225	CD31 Tubule Staining Kit	Mouse anti-human CD31 primary antibody, goat anti-mouse IgG1-Alk. Phos. secondary antibody and BCIP/NBT tablets
	ZHA-1970	Angiogenesis Growth Medium Package	Angiogenesis basal medium, growth supplement and antibiotic/antimycotic supplement
Software	ZHA-5000	AngioSys 2.0	Cellworks Image Analysis Software (full licence)
	ZHA-5000D	AngioSys 2.0 (demonstration version)	Cellworks Image Analysis Software (free 21-day demonstration licence)

10. References

1. Klagsbrun, M. & D'Amore, P.A. (1991) *Ann. Rev. Physiol.* **53**; 217
2. Findlay, J.K. (1986) *J. Endocrinol.* **111**; 357
3. Folkman, J. (1972) *Ann. Surg.* **175**; 409
4. Davis, M.D. (1988) *Diabetes/Metab. Rev.* **4**; 291
5. Folkman, J. (1985) *Adv. Cancer Res.* **43**; 175
6. Folkman, J. & Klagsbrun, M. (1987) *Science* **235**; 442
7. Ausprunk, D.H. & Folkman, J. (1977) *Microvasc. Res.* **14**; 53
8. Sholley, M.M. *et al* (1984) *Lab. Invest.* **51**; 624
9. Vu *et al* (1985) *Lab. Invest.* **53**; 499
10. Gimbrone, M.A. *et al* (1972) *J. Exp. Med.* **136**; 261
11. Connolly, D.T. *et al* (1989) *J. Clin. Invest.* **84**; 1470
12. Rupnick, M.A. *et al* (1988) *Lab. Invest.* **59**; 363
13. Passaniti, A. *et al* (1992) *Lab. Invest.* **67**; 519
14. Jones *et al* (1998) *Biochem Biophys Res Commun* **249(1)**; 118

11. Technical Support

For assistance please contact:

Caltag Medsystems Ltd
Whiteleaf Business Centre
11 Little Balmer
Buckingham
MK18 1TF
UK

Tel: +44 (0)1280 827460

Fax: +44 (0)1280 827466

E-mail: techsupport@caltagmedsystems.co.uk (Technical Support)

office@caltagmedsystems.co.uk (Sales)

Web: www.cellworks.co.uk



Tel: +44 (0)1280 827460

Fax: +44 (0)1280 827466

E-mail: techsupport@caltagmedsystems.co.uk
office@caltagmedsystems.co.uk

Web: www.cellworks.co.uk

Caltag Medsystems Ltd, Whiteleaf Business Centre,
11 Little Balmer, Buckingham, MK18 1TF, UK



A Caltag Medsystems Company

Blank page